

SOLUTION CHANGE

BFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 12/01/2021

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:07
INTERNAL STANDARD	VERIFIED	03:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:08
BLANK TEST	0.00	03:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:09
BLANK TEST	0.00	03:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:10
BLANK TEST	0.00	03:11

