

BFO

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 08/26/2021

OPERATOR NAME:  
ROBERT S CORNISH  
Permit #: 48068  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:13
INTERNAL STANDARD	VERIFIED	01:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:13
BLANK TEST	0.00	01:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:14
BLANK TEST	0.00	01:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:15
BLANK TEST	0.00	01:16

