

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 11/11/2021

OPERATOR NAME:  
JAMES P QUINN  
Permit #: 49340  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:19
INTERNAL STANDARD	VERIFIED	02:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:19
BLANK TEST	0.00	02:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:20
BLANK TEST	0.00	02:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:22
BLANK TEST	0.00	02:23

BFO DataMaster  
weekly test  
11/11/21

