

SUPERVISOR TEST

BFD

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 11/17/2021

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:29
INTERNAL STANDARD	VERIFIED	18:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:30
BLANK TEST	0.00	18:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:31
BLANK TEST	0.00	18:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:32
BLANK TEST	0.00	18:33

