

BFO

SOLUTION CHANGE

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 11/01/2022

OPERATOR NAME:  
LEONARD F DILLENBECK  
Permit #: 47621  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:10
INTERNAL STANDARD	VERIFIED	02:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:11
BLANK TEST	0.00	02:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:12
BLANK TEST	0.00	02:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:13
BLANK TEST	0.00	02:14

