

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 03/24/2022

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:07
INTERNAL STANDARD	VERIFIED	00:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:07
BLANK TEST	0.00	00:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:09
BLANK TEST	0.00	00:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:10
BLANK TEST	0.00	00:11

