

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 05/04/2022

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:30
INTERNAL STANDARD	VERIFIED	18:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:31
BLANK TEST	0.00	18:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:32
BLANK TEST	0.00	18:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:33
BLANK TEST	0.00	18:34

