

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 06/08/2022

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:10
INTERNAL STANDARD	VERIFIED	01:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:11
BLANK TEST	0.00	01:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:12
BLANK TEST	0.00	01:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:13
BLANK TEST	0.00	01:14

