

BFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 06/22/2022

OPERATOR NAME:
LEONARD F DILLENBECK
Permit #: 47621
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22100

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 17:53 |
| INTERNAL STANDARD | VERIFIED | 17:53 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 17:54 |
| BLANK TEST | 0.00 | 17:55 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 17:55 |
| BLANK TEST | 0.00 | 17:56 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 17:56 |
| BLANK TEST | 0.00 | 17:57 |

