

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

BFO

Date: 06/29/2022

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:55
INTERNAL STANDARD	VERIFIED	17:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:56
BLANK TEST	0.00	17:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:57
BLANK TEST	0.00	17:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:58
BLANK TEST	0.00	17:59

