

BFO

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 07/27/2022

OPERATOR NAME:  
LEONARD F DILLENBECK  
Permit #: 47621  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:52
INTERNAL STANDARD	VERIFIED	01:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:52
BLANK TEST	0.00	01:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:53
BLANK TEST	0.00	01:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:55
BLANK TEST	0.00	01:56

