

BFO

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 11/16/2022

OPERATOR NAME:  
LEONARD F DILLENBECK  
Permit #: 47621  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:21
INTERNAL STANDARD	VERIFIED	03:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:21
BLANK TEST	0.00	03:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:23
BLANK TEST	0.00	03:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:24
BLANK TEST	0.00	03:25

