

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 12/28/2022

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:36
INTERNAL STANDARD	VERIFIED	00:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:37
BLANK TEST	0.00	00:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:38
BLANK TEST	0.00	00:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:39
BLANK TEST	0.00	00:40

