

SOLUTION CHANGE

BFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 04/01/2023

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:47
INTERNAL STANDARD	VERIFIED	20:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:48
BLANK TEST	0.00	20:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:49
BLANK TEST	0.00	20:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:51
BLANK TEST	0.00	20:52

