

SUPERVISOR TEST

BFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 04/05/2023

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:53
INTERNAL STANDARD	VERIFIED	18:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:54
BLANK TEST	0.00	18:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:56
BLANK TEST	0.00	18:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:57
BLANK TEST	0.00	18:58

