

SUPERVISOR TEST

 ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 141906

Date: 04/19/2023

BFO

OPERATOR NAME:
 STEVEN P EILERS
 Permit #: 23030
 Operator Agency/Dept: ONEIDA CO. S.O.
 Reference Standard #: 49321

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:04
INTERNAL STANDARD	VERIFIED	01:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:05
BLANK TEST	0.00	01:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:06
BLANK TEST	0.00	01:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:07
BLANK TEST	0.00	01:08

