

SUPERVISOR TEST

BFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 06/21/2023

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:25
INTERNAL STANDARD	VERIFIED	02:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:25
BLANK TEST	0.00	02:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:26
BLANK TEST	0.00	02:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:27
BLANK TEST	0.00	02:28

