

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

BFO

Date: 07/12/2023

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:03
INTERNAL STANDARD	VERIFIED	18:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:04
BLANK TEST	0.00	18:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:05
BLANK TEST	0.00	18:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:06
BLANK TEST	0.00	18:07

