

SUPERVISOR TEST

BFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 09/06/2023

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:01
INTERNAL STANDARD	VERIFIED	18:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:01
BLANK TEST	0.00	18:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:02
BLANK TEST	0.00	18:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:03
BLANK TEST	0.00	18:04

