

SUPERVISOR TEST

BFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 11/22/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:06
INTERNAL STANDARD	VERIFIED	00:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:06
BLANK TEST	0.00	00:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:08
BLANK TEST	0.00	00:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:09
BLANK TEST	0.00	00:10

