

SOLUTION CHANGE

BFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 03/22/2024

OPERATOR NAME:  
JOHN P CUSACK  
Permit #: 35563  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:50
INTERNAL STANDARD	VERIFIED	00:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:51
BLANK TEST	0.00	00:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:52
BLANK TEST	0.00	00:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:53
BLANK TEST	0.00	00:54

