

SUPERVISOR TEST

BFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 01/24/2024

OPERATOR NAME:  
JOHN P CUSACK  
Permit #: 35563  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:37
INTERNAL STANDARD	VERIFIED	19:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:37
BLANK TEST	0.00	19:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:39
BLANK TEST	0.00	19:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:40
BLANK TEST	0.00	19:41

