

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 09/15/2021

OPERATOR NAME:
ROBERT S CORNISH
Permit #: 48068
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:03
INTERNAL STANDARD	VERIFIED	06:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:03
BLANK TEST	0.00	06:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:04
BLANK TEST	0.00	06:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:06
BLANK TEST	0.00	06:07

