

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 10/04/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:47
INTERNAL STANDARD	VERIFIED	20:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:48
BLANK TEST	0.00	20:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:49
BLANK TEST	0.00	20:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:50
BLANK TEST	0.00	20:51

