

SUPERVISOR TEST

LEB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 12/22/2021

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:20
INTERNAL STANDARD	VERIFIED	18:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:20
BLANK TEST	0.00	18:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:22
BLANK TEST	0.00	18:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:23
BLANK TEST	0.00	18:24

