

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LEB

Date: 12/29/2021

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:55
INTERNAL STANDARD	VERIFIED	23:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:55
BLANK TEST	0.00	23:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:56
BLANK TEST	0.00	23:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:58
BLANK TEST	0.00	23:59

