

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 06/01/2022

LEB

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:03
INTERNAL STANDARD	VERIFIED	05:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:04
BLANK TEST	0.00	05:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:05
BLANK TEST	0.00	05:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:06
BLANK TEST	0.00	05:07

