

**SOLUTION CHANGE**

-----  
 ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 113606  
 -----

Date: 08/01/2022

OPERATOR NAME:  
 URIAH J PORTNER  
 Permit #: 50824  
 Operator Agency/Dept: ONEIDA CO. SO  
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:19
INTERNAL STANDARD	VERIFIED	18:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:19
BLANK TEST	0.00	18:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:20
BLANK TEST	0.00	18:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:21
BLANK TEST	0.00	18:22

