

SUPERVISOR TEST

 ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113606

Date: 07/27/2022

OPERATOR NAME:
 JAY V SORRENTINO
 Permit #: 38962
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:54
INTERNAL STANDARD	VERIFIED	05:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:55
BLANK TEST	0.00	05:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:56
BLANK TEST	0.00	05:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:57
BLANK TEST	0.00	05:58

