

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

*LER*

Date: 08/03/2022

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:10
INTERNAL STANDARD	VERIFIED	22:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:10
BLANK TEST	0.00	22:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:12
BLANK TEST	0.00	22:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:13
BLANK TEST	0.00	22:14

