

SUPERVISOR TEST

LEB

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113606

Date: 08/10/2022

OPERATOR NAME:
 AARON M ALSHAMAN
 Permit #: 50807
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:10
INTERNAL STANDARD	VERIFIED	18:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:10
BLANK TEST	0.00	18:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:11
BLANK TEST	0.00	18:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:13
BLANK TEST	0.00	18:14

