

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LEB

Date: 09/14/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:50
INTERNAL STANDARD	VERIFIED	23:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:51
BLANK TEST	0.00	23:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:52
BLANK TEST	0.00	23:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:53
BLANK TEST	0.00	23:54

