

**SUPERVISOR TEST**

-----  
 ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 113606  
 -----

Date: 10/05/2022

*LER*

OPERATOR NAME:  
 AARON M ALSHAMAN  
 Permit #: 50807  
 Operator Agency/Dept: ONEIDA CO. SO  
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:10
INTERNAL STANDARD	VERIFIED	22:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:10
BLANK TEST	0.00	22:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:11
BLANK TEST	0.00	22:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:13
BLANK TEST	0.00	22:13

