

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 11/16/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:19
INTERNAL STANDARD	VERIFIED	03:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:19
BLANK TEST	0.00	03:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:20
BLANK TEST	0.00	03:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:22
BLANK TEST	0.00	03:22

LEAD

