

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LEB

Date: 12/14/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:58
INTERNAL STANDARD	VERIFIED	00:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:59
BLANK TEST	0.00	01:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:00
BLANK TEST	0.00	01:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:01
BLANK TEST	0.00	01:02

