

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 12/28/2022

LEB

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:48
INTERNAL STANDARD	VERIFIED	03:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:48
BLANK TEST	0.00	03:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:49
BLANK TEST	0.00	03:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:51
BLANK TEST	0.00	03:51

