

SUPERVISOR TEST

LAB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 01/26/2022

OPERATOR NAME:  
ZACHARY S PRZYBYLA  
Permit #: 49339  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:21
INTERNAL STANDARD	VERIFIED	19:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:21
BLANK TEST	0.00	19:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:23
BLANK TEST	0.00	19:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:24
BLANK TEST	0.00	19:25

