

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 02/09/2022

OPERATOR NAME:
ROBERT S CORNISH
Permit #: 48068
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:57
INTERNAL STANDARD	VERIFIED	23:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:57
BLANK TEST	0.00	23:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:59
BLANK TEST	0.00	00:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:00
BLANK TEST	0.00	00:01

