

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 03/16/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:39
INTERNAL STANDARD	VERIFIED	18:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:39
BLANK TEST	0.00	18:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:40
BLANK TEST	0.00	18:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:41
BLANK TEST	0.00	18:42

