

LEB

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 04/06/2022

OPERATOR NAME:  
ROBERT S CORNISH  
Permit #: 48068  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:05
INTERNAL STANDARD	VERIFIED	19:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:06
BLANK TEST	0.00	19:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:07
BLANK TEST	0.00	19:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:08
BLANK TEST	0.00	19:09

