

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

LEB

Date: 05/25/2022

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:10
INTERNAL STANDARD	VERIFIED	21:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:10
BLANK TEST	0.00	21:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:11
BLANK TEST	0.00	21:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:12
BLANK TEST	0.00	21:13

