

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 06/08/2022

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:56
INTERNAL STANDARD	VERIFIED	03:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:56
BLANK TEST	0.00	03:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:58
BLANK TEST	0.00	03:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:59
BLANK TEST	0.00	04:00

