

SOLUTION CHANGE

 ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113606

Date: 05/03/2023

OPERATOR NAME:
 BAILEY J DIXON
 Permit #: 50814
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 00:33 |
| INTERNAL STANDARD | VERIFIED | 00:34 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 00:34 |
| BLANK TEST | 0.00 | 00:35 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 00:35 |
| BLANK TEST | 0.00 | 00:36 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 00:36 |
| BLANK TEST | 0.00 | 00:37 |

