

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 06/01/2023

LEB

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:41
INTERNAL STANDARD	VERIFIED	02:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:41
BLANK TEST	0.00	02:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:43
BLANK TEST	0.00	02:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:44
BLANK TEST	0.00	02:45

