

SOLUTION CHANGE

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113606

LEB

Date: 10/01/2023

OPERATOR NAME:
 JOHN P CUSACK
 Permit #: 35563
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:08
INTERNAL STANDARD	VERIFIED	05:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:09
BLANK TEST	0.00	05:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:10
BLANK TEST	0.00	05:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:11
BLANK TEST	0.00	05:12

