

SOLUTION CHANGE

LEB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 12/01/2023

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:16
INTERNAL STANDARD	VERIFIED	05:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:16
BLANK TEST	0.00	05:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:18
BLANK TEST	0.00	05:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:19
BLANK TEST	0.00	05:20

