

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 05/31/2023

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:57
INTERNAL STANDARD	VERIFIED	18:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:57
BLANK TEST	0.00	18:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:58
BLANK TEST	0.00	18:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:59
BLANK TEST	0.00	19:00

