

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 07/12/2023

LEB

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:11
INTERNAL STANDARD	VERIFIED	18:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:12
BLANK TEST	0.00	18:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:13
BLANK TEST	0.00	18:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:14
BLANK TEST	0.00	18:15

