

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

LEB

Date: 08/16/2023

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:07
INTERNAL STANDARD	VERIFIED	05:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:08
BLANK TEST	0.00	05:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:09
BLANK TEST	0.00	05:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:10
BLANK TEST	0.00	05:11

