

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 08/23/2023

LEB

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:48
INTERNAL STANDARD	VERIFIED	20:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:49
BLANK TEST	0.00	20:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:50
BLANK TEST	0.00	20:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:51
BLANK TEST	0.00	20:52

